



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Center - WO66-G609
Silver Spring, MD 20993-0002

Karl Storz Endoscopy-America, Inc.
Ms. Monika Campbell
Senior Regulatory Affairs Specialist
600 Corporate Pointe, 5th Floor
Culver City, CA 90230

JUL 27 2015

Re: K070827
Trade/Device Name: SCB Interface Control
Regulation Number: 21 CFR 876.1500
Regulation Name: Endoscope and accessories
Regulatory Class: II
Product Code: ODA
Dated (Date on orig SE ltr): March 22, 2007
Received (Date on orig SE ltr): March 26, 2007

Dear Ms. Campbell,

This letter corrects our substantially equivalent letter of May 22, 2007.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be

found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

Benjamin R. Fisher -S

Benjamin R. Fisher, Ph.D.
Director
Division of Reproductive, Gastro-Renal,
and Urological Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K070827

Device Name: SCB Interface Control

Indications for Use: The KSEA SCB Interface Control is a remote control unit for RS-232 compatible medical/non-medical equipment to the Karl Storz SCB Control Computer (Karl Storz Communication Bus), to relay functions and controls on the SCB monitor.

Prescription Use X
(Part 21 CFR 801 Subpart D)

~~AND/OR~~

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Nancy Brogdon
(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K070827

Page 9 of 172

(Posted November 13, 2003)

K070827
PG 1 of 1

510(k) SUMMARY OF SAFETY AND EFFECTIVENESS

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of the Safe Medical Devices Act (SMDA) of 1990 and 21 CFR 807.92. All data included in this document is accurate and complete to the best of KSEA's knowledge.

Applicant: Karl Storz Endoscopy - America, Inc.
600 Corporate Pointe Drive
Culver City, CA 90230
(310) 558-1500

MAY 22 2007

Contact: Monika Campbell
Senior Regulatory Affairs Specialist

Device Identification: Common Name:
Remote Control

Trade Name: (optional)
Karl Storz SCB Interface Control

Indication: The KSEA SCB Interface Control is a remote control unit for RS-232 compatible medical/non-medical equipment to the Karl Storz SCB Control Computer (Karl Storz Communication Bus), to relay functions and controls on the SCB monitor.

Device Description: The Karl Storz SCB Interface Control is a remote control unit.

Substantial Equivalence: The Karl Storz SCB Interface Control is substantially equivalent to the predicate device since the basic features and intended uses are the same. The minor differences between the Karl Storz Interface Control and the predicate device raise no new issues of safety and effectiveness, as these differences have no effect on the performance, function or general intended use of these devices.

Signed: 

Monika Campbell
Senior Regulatory Affairs Specialist